

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 30 °C             | 30.213 °C         | 30.7 °C   | 0.487 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.059 %HR        | 29.9 %HR  | -0.159 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 50 %HR            | 50.243 %HR        | 47.4 %HR  | -2.843 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.448 %HR        | 73.4 %HR  | 2.952 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |