

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.030 °C	0.1 °C	0.070 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.943 °C	5.0 °C	0.057 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.408 °C	10.3 °C	-0.108 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	29.1 %HR	-1.061 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.171 %HR	52.6 %HR	2.429 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.550 %HR	73.5 %HR	2.950 %HR	± 3.0 %HR	± 1.9E+00 %HR