

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
12 °C	12.213 °C	12.3 °C	0.087 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.255 °C	27.4 °C	0.145 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.095 °C	42.1 °C	0.005 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.349 %HR	53.3 %HR	0.951 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	72.288 %HR	72.4 %HR	0.112 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.022 %HR	94.7 %HR	2.678 %HR	± 3.0 %HR	± 1.9E+00 %HR