

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
30.17353 µl	# 5 µl	-0.17353 µl	± 0.125 µl	±	5.1E-02 µl
100.13401 µl	100 µl	-0.13401 µl	± 0.5 µl	±	5.1E-02 µl
150.25615 µl	150 µl	-0.25615 µl	± 0.5 µl	±	5.2E-02 µl