

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
30.17353 µl	# 5 µl	-0.17353 µl	± 0.5 µl	±	5.1E-02 µl
100.13401 µl	100 µl	-0.13401 µl	± 0.8 µl	±	5.1E-02 µl
150.25615 µl	150 µl	-0.25615 µl	± 1.5 µl	±	5.2E-02 µl