

Patrón <i>(Standard)</i>		UBP <i>(UUT)</i>		Error <i>(Error)</i>		E.M.P <i>(Tolerance)</i>		Incertidumbre <i>(Uncertainty)</i>	
0.0625	in	0.0626	in	0.0001	in	± 0.0003	in	±	5.9E-05 in
0.100	in	0.1001	in	0.0001	in	± 0.0003	in	±	6.0E-05 in
0.125	in	0.1251	in	0.0001	in	± 0.0003	in	±	6.1E-05 in
0.200	in	0.2001	in	0.0001	in	± 0.0003	in	±	6.3E-05 in
0.250	in	0.2501	in	0.0001	in	± 0.0003	in	±	6.4E-05 in
0.300	in	0.3001	in	0.0001	in	± 0.0003	in	±	6.5E-05 in
0.400	in	0.4001	in	0.0001	in	± 0.0003	in	±	6.5E-05 in
0.500	in	0.5001	in	0.0001	in	± 0.0003	in	±	6.5E-05 in
0.600	in	0.6001	in	0.0001	in	± 0.0003	in	±	6.5E-05 in
0.800	in	0.8001	in	0.0001	in	± 0.0003	in	±	6.5E-05 in
1.000	in	1.0001	in	0.0001	in	± 0.0003	in	±	6.5E-05 in