

Patrón <i>(Standard)</i>		UBP <i>(UUT)</i>		Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.0625	in	1.0625	in	0.0000 in	± 0.0003 in	±	5.9E-05 in
1.100	in	1.1030	in	0.0030 in	± 0.0003 in	±	6.0E-05 in
1.125	in	1.1280	in	0.0030 in	± 0.0003 in	±	6.1E-05 in
1.200	in	1.2060	in	0.0060 in	± 0.0003 in	±	6.3E-05 in
1.250	in	1.2560	in	0.0060 in	± 0.0003 in	±	6.4E-05 in
1.300	in	1.3100	in	0.0100 in	± 0.0003 in	±	6.5E-05 in
1.400	in	1.4100	in	0.0100 in	± 0.0003 in	±	6.5E-05 in
1.500	in	1.5100	in	0.0100 in	± 0.0003 in	±	6.5E-05 in
1.600	in	1.6100	in	0.0100 in	± 0.0003 in	±	6.5E-05 in
1.800	in	1.8100	in	0.0100 in	± 0.0003 in	±	6.5E-05 in
2.000	in	2.0200	in	0.0200 in	± 0.0003 in	±	6.5E-05 in