Patrón (Standard)		UBP (UUT)		Error (Error)	E.M.P (Tolerance)	Incertidum	Incertidumbre (Uncertainty)	
2.0625	in	2.0640	in	0.0015 in	± 0.0003 in	±	5.9E-05 in	
2.100	in	2.1015	in	0.0015 in	$\pm 0.0003 in$	±	6.0E-05 in	
2.125	in	2.1265	in	0.0015 in	$\pm 0.0003 in$	±	6.1E-05 in	
2.200	in	2.2015	in	0.0015 in	$\pm 0.0003 in$	±	6.3E-05 in	
2.250	in	2.2515	in	0.0015 in	$\pm 0.0003 in$	±	6.4E-05 in	
2.300	in	2.3015	in	0.0015 in	$\pm 0.0003 in$	±	6.5E-05 in	
2.400	in	2.4015	in	0.0015 in	$\pm 0.0003 in$	±	6.5E-05 in	
2.500	in	2.5015	in	0.0015 in	$\pm 0.0003 in$	±	6.5E-05 in	
2.600	in	2.6015	in	0.0015 in	± 0.0003 in	±	6.5E-05 in	
2.800	in	2.8015	in	0.0015 in	± 0.0003 in	±	6.5E-05 in	
3.000	in	3.0015	in	0.0015 in	$\pm 0.0003 in$	±	6.5E-05 in	