

Patrón <i>(Standard)</i>		UBP <i>(UUT)</i>		Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
3.0625	in	3.0626	in	0.0001 in	± 0.0003 in	±	5.9E-05 in
3.100	in	3.1001	in	0.0001 in	± 0.0003 in	±	6.0E-05 in
3.125	in	3.1251	in	0.0001 in	± 0.0003 in	±	6.1E-05 in
3.200	in	3.2001	in	0.0001 in	± 0.0003 in	±	6.3E-05 in
3.250	in	3.2501	in	0.0001 in	± 0.0003 in	±	6.4E-05 in
3.300	in	3.3001	in	0.0001 in	± 0.0003 in	±	6.5E-05 in
3.400	in	3.4001	in	0.0001 in	± 0.0003 in	±	6.5E-05 in
3.500	in	3.5001	in	0.0001 in	± 0.0003 in	±	6.5E-05 in
3.600	in	3.6001	in	0.0001 in	± 0.0003 in	±	6.5E-05 in
3.800	in	3.8010	in	0.0010 in	± 0.0003 in	±	6.5E-05 in
4.000	in	4.0010	in	0.0010 in	± 0.0003 in	±	6.5E-05 in