

Patrón <i>(Standard)</i>		UBP <i>(UUT)</i>		Error <i>(Error)</i>		E.M.P <i>(Tolerance)</i>		Incertidumbre <i>(Uncertainty)</i>	
4.0625	in	4.0626	in	0.0001	in	± 0.0003	in	±	5.9E-05 in
4.100	in	4.1001	in	0.0001	in	± 0.0003	in	±	6.0E-05 in
4.125	in	4.1251	in	0.0001	in	± 0.0003	in	±	6.1E-05 in
4.200	in	4.2001	in	0.0001	in	± 0.0003	in	±	6.3E-05 in
4.250	in	4.2501	in	0.0001	in	± 0.0003	in	±	6.4E-05 in
4.300	in	4.3001	in	0.0001	in	± 0.0003	in	±	6.5E-05 in
4.400	in	4.4001	in	0.0001	in	± 0.0003	in	±	6.5E-05 in
4.500	in	4.5001	in	0.0001	in	± 0.0003	in	±	6.5E-05 in
4.600	in	4.6001	in	0.0001	in	± 0.0003	in	±	6.5E-05 in
4.800	in	4.8010	in	0.0010	in	± 0.0003	in	±	6.5E-05 in
5.000	in	5.0010	in	0.0010	in	± 0.0003	in	±	6.5E-05 in