

Patrón <i>(Standard)</i>		UBP <i>(UUT)</i>		Error <i>(Error)</i>		E.M.P <i>(Tolerance)</i>		Incertidumbre <i>(Uncertainty)</i>	
5.0625	in	5.0626	in	0.0001	in	± 0.0003	in	±	5.9E-05 in
5.100	in	5.1001	in	0.0001	in	± 0.0003	in	±	6.0E-05 in
5.125	in	5.1251	in	0.0001	in	± 0.0003	in	±	6.1E-05 in
5.200	in	5.2001	in	0.0001	in	± 0.0003	in	±	6.3E-05 in
5.250	in	5.2501	in	0.0001	in	± 0.0003	in	±	6.4E-05 in
5.300	in	5.3001	in	0.0001	in	± 0.0003	in	±	6.5E-05 in
5.400	in	5.4001	in	0.0001	in	± 0.0003	in	±	6.5E-05 in
5.500	in	5.5001	in	0.0001	in	± 0.0003	in	±	6.5E-05 in
5.600	in	5.6001	in	0.0001	in	± 0.0003	in	±	6.5E-05 in
5.800	in	5.8010	in	0.0010	in	± 0.0003	in	±	6.5E-05 in
6.000	in	6.0010	in	0.0010	in	± 0.0003	in	±	6.5E-05 in