

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.866 °C	30.5 °C	0.634 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.261 %HR	30 %HR	0.739 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.128 %HR	51 %HR	1.872 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.364 %HR	72 %HR	2.636 %HR	± 3.0 %HR	± 1.9E+00 %HR