

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|------------|
| 50.0 mmHg                | 50 mmHg          | 0.0 mmHg             | ± 3 mmHg                    | ±                                  | #REF! mmHg |
| 100.0 mmHg               | 100 mmHg         | 0.0 mmHg             | ± 3 mmHg                    | ±                                  | #REF! mmHg |
| 150.0 mmHg               | 150 mmHg         | 0.0 mmHg             | ± 3 mmHg                    | ±                                  | #REF! mmHg |
| 200.0 mmHg               | 200 mmHg         | 0.0 mmHg             | ± 3 mmHg                    | ±                                  | #REF! mmHg |
| 250.0 mmHg               | 250 mmHg         | 0.0 mmHg             | ± 3 mmHg                    | ±                                  | #REF! mmHg |
| 300.0 mmHg               | 300 mmHg         | 0.0 mmHg             | ± 3 mmHg                    | ±                                  | #REF! mmHg |