

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
50.00 mmHg	50 mmHg	0.00 mmHg	± 3 mmHg	±	6.6E-01 mmHg
100.00 mmHg	100 mmHg	0.00 mmHg	± 3 mmHg	±	6.5E-01 mmHg
149.99 mmHg	150 mmHg	0.01 mmHg	± 3 mmHg	±	6.6E-01 mmHg
200.00 mmHg	200 mmHg	0.00 mmHg	± 3 mmHg	±	6.8E-01 mmHg
249.99 mmHg	250 mmHg	0.01 mmHg	± 3 mmHg	±	7.5E-01 mmHg
300.01 mmHg	300 mmHg	-0.01 mmHg	± 3 mmHg	±	7.5E-01 mmHg