

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |              |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|--------------|
| 50.00 mmHg               | 50 mmHg          | 0.00 mmHg            | ± 3 mmHg                    | ±                                  | 6.6E-01 mmHg |
| 100.00 mmHg              | 100 mmHg         | 0.00 mmHg            | ± 3 mmHg                    | ±                                  | 6.5E-01 mmHg |
| 149.99 mmHg              | 150 mmHg         | 0.01 mmHg            | ± 3 mmHg                    | ±                                  | 6.6E-01 mmHg |
| 200.00 mmHg              | 200 mmHg         | 0.00 mmHg            | ± 3 mmHg                    | ±                                  | 6.8E-01 mmHg |
| 250.00 mmHg              | 250 mmHg         | 0.00 mmHg            | ± 3 mmHg                    | ±                                  | 7.5E-01 mmHg |
| 299.99 mmHg              | 300 mmHg         | 0.01 mmHg            | ± 3 mmHg                    | ±                                  | 7.5E-01 mmHg |