

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	20.060 °C	19.5 °C	-0.560 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.584 °C	31.0 °C	0.416 °C	± 1.0 °C	± 4.0E-01 °C
40 %HR	39.302 %HR	41 %HR	1.698 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	58.147 %HR	61 %HR	2.853 %HR	± 3.0 %HR	± 1.9E+00 %HR