

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.012 °C	30.3 °C	0.288 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.077 %HR	30.1 %HR	1.023 %HR	± 2.5 %HR	± 1.9E+00 %HR
50 %HR	49.507 %HR	50.2 %HR	0.693 %HR	± 2.5 %HR	± 1.9E+00 %HR
70 %HR	70.437 %HR	71.3 %HR	0.863 %HR	± 2.5 %HR	± 1.9E+00 %HR