

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.882 °C	30.1 °C	0.218 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.781 %HR	30.0 %HR	1.219 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	48.750 %HR	49.9 %HR	1.150 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.364 %HR	70.1 %HR	0.736 %HR	± 3.0 %HR	± 1.9E+00 %HR