

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>		Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
3600.057 s	3600.0 s	-0.057 s	-0.0016%	± 0.01%	±	2.9E-01 s
3600.066 s	3600.0 s	-0.066 s	-0.0018%	± 0.01%	±	2.9E-01 s
3600.081 s	3600.0 s	-0.081 s	-0.0022%	± 0.01%	±	2.9E-01 s