

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	22.050 °C	22.5 °C	0.450 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.259 °C	25.7 °C	0.441 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	30.855 °C	31.2 °C	0.345 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	47.216 %HR	48.7 %HR	1.484 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.124 %HR	57.6 %HR	0.476 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	66.756 %HR	67.4 %HR	0.644 %HR	± 3.0 %HR	± 1.9E+00 %HR