

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 22 °C                      | 22.050 °C                  | 22.4 °C            | 0.350 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 25 °C                      | 25.259 °C                  | 25.4 °C            | 0.141 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 31 °C                      | 30.855 °C                  | 31.1 °C            | 0.245 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 48 %HR                     | 47.216 %HR                 | 48.4 %HR           | 1.184 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 57 %HR                     | 57.124 %HR                 | 57.6 %HR           | 0.476 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 67 %HR                     | 66.756 %HR                 | 67.5 %HR           | 0.744 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |