

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
2 °C	2.035 °C	2.2 °C	0.165 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.973 °C	5.2 °C	0.227 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.941 °C	8.1 °C	0.159 °C	± 0.5 °C	± 4.0E-01 °C