

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
# -30 °C	-30.304 °C	-30.0 °C	0.304 °C	± 0.45 °C	± 4.0E-01 °C
-25 °C	-25.491 °C	-26.0 °C	-0.509 °C	± 0.45 °C	± 4.0E-01 °C
-15 °C	-15.563 °C	-15.5 °C	0.063 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.449 °C	-11.0 °C	-0.551 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	1.985 °C	1.5 °C	-0.485 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.042 °C	8.0 °C	-0.042 °C	± 0.5 °C	± 4.0E-01 °C