

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
# -30 °C	-30.304 °C	-30.6 °C	-0.296 °C	± 0.5 °C	± 4.0E-01 °C
-25 °C	-25.491 °C	-25.2 °C	0.291 °C	± 0.5 °C	± 4.0E-01 °C
-15 °C	-15.563 °C	-15.4 °C	0.163 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.449 °C	-10.6 °C	-0.151 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	1.985 °C	2.0 °C	0.015 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.042 °C	8.0 °C	-0.042 °C	± 0.5 °C	± 4.0E-01 °C