

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.985 °C	1.5 °C	-0.485 °C	± 0.45 °C	± 4.0E-01 °C
8 °C	8.042 °C	8.5 °C	0.458 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.408 °C	10.5 °C	0.092 °C	± 0.5 °C	± 4.0E-01 °C