

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	1.985 °C	1.5 °C	-0.485 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.042 °C	8.5 °C	0.458 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.408 °C	10.5 °C	0.092 °C	± 0.5 °C	± 4.0E-01 °C