

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
# -30 °C	-30.304 °C	-30.5 °C	-0.196 °C	± 0.5 °C	± 4.0E-01 °C
-25 °C	-25.491 °C	-25.5 °C	-0.009 °C	± 0.5 °C	± 4.0E-01 °C
-15 °C	-15.563 °C	-16.0 °C	-0.437 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.449 °C	-10.5 °C	-0.051 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	1.985 °C	1.5 °C	-0.485 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.042 °C	8.0 °C	-0.042 °C	± 0.5 °C	± 4.0E-01 °C