

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.522 °C	15.4 °C	-0.122 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.269 °C	25.0 °C	-0.269 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.213 °C	30.0 °C	-0.213 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	60.018 %HR	62.2 %HR	2.182 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	65.540 %HR	67.0 %HR	1.460 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.550 %HR	70.3 %HR	-0.250 %HR	± 3.0 %HR	± 1.9E+00 %HR