

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.522 °C	15.3 °C	-0.222 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.269 °C	25.4 °C	0.131 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.213 °C	30.4 °C	0.187 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	60.018 %HR	62.5 %HR	2.482 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	65.540 %HR	68.4 %HR	2.860 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.550 %HR	71.4 %HR	0.850 %HR	± 3.0 %HR	± 1.9E+00 %HR