

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.522 °C	15.7 °C	0.178 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.269 °C	25.2 °C	-0.069 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.213 °C	30.0 °C	-0.213 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	60.018 %HR	61.0 %HR	0.982 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	65.540 %HR	66.1 %HR	0.560 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.550 %HR	68.1 %HR	-2.450 %HR	± 3.0 %HR	± 1.9E+00 %HR