

Certificado No. (Certificate #):		60448		Equipo (Instrument):		Medidor de Vibración			
Fecha de Calibración (Calibration Date):		2025-08-25		Marca (Brand):		Monarch Instrument			
Patrón (Standard)		UBP (UUT)		Error (Error)		Exactitud (Accuracy)		Incertidumbre (Uncertainty)	
7 Hz	0.28 g-rms	0.29 g-rms	0.01 g-rms	± 0.0085 g-rms	±	5.8E-02	g-rms		
10 Hz	0.57 g-rms	0.55 g-rms	-0.02 g-rms	± 0.0170 g-rms	±	5.9E-02	g-rms		
50 Hz	0.71 g-rms	0.69 g-rms	-0.02 g-rms	± 0.0212 g-rms	±	5.9E-02	g-rms		
100 Hz	0.71 g-rms	0.69 g-rms	-0.02 g-rms	± 0.0212 g-rms	±	6.0E-02	g-rms		
1000 Hz	0.71 g-rms	0.69 g-rms	-0.02 g-rms	± 0.0212 g-rms	±	6.0E-02	g-rms		
5000 Hz	0.71 g-rms	0.68 g-rms	-0.03 g-rms	± 0.0212 g-rms	±	6.2E-02	g-rms		