

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.000 mm	1.002 mm	2.00 μm	± 2 μm	±	1.2E+00 μm
1.500 mm	1.502 mm	2.00 μm	± 2 μm	±	1.2E+00 μm
2.000 mm	2.002 mm	2.00 μm	± 2 μm	±	1.2E+00 μm
3.000 mm	3.002 mm	2.00 μm	± 2 μm	±	1.2E+00 μm
4.000 mm	4.002 mm	2.00 μm	± 2 μm	±	1.2E+00 μm
5.000 mm	5.004 mm	4.00 μm	± 2 μm	±	1.2E+00 μm