

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.912 °C	29.9 °C	-0.012 °C	± 0.6 °C	± 4.0E-01 °C
30 %HR	29.190 %HR	27 %HR	-2.090 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.233 %HR	48 %HR	-2.133 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.233 %HR	73 %HR	-2.433 %HR	± 3.0 %HR	± 1.9E+00 %HR