

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.282 °C	15.5 °C	0.218 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.193 °C	30.3 °C	0.107 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.146 °C	35.3 °C	0.154 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	39.241 %HR	40 %HR	1.059 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.693 %HR	52 %HR	1.407 %HR	± 3.0 %HR	± 1.9E+00 %HR
90 %HR	90.008 %HR	92 %HR	1.692 %HR	± 3.0 %HR	± 1.9E+00 %HR