

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0625 in	0.062 in	-0.0005 in	± 40 in	±	7.0E-02 in
0.100 in	0.101 in	0.0010 in	± 40 in	±	7.6E-02 in
0.125 in	0.124 in	-0.0010 in	± 40 in	±	9.4E-02 in
0.200 in	0.198 in	-0.0020 in	± 40 in	±	1.0E-01 in
0.250 in	0.248 in	-0.0020 in	± 40 in	±	7.0E-02 in
0.300 in	0.298 in	-0.0020 in	± 40 in	±	1.2E-01 in
0.500 in	0.499 in	-0.0010 in	± 40 in	±	9.9E-02 in