

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0625 in	0.062 in	-0.0005 in	± 0.001 in	±	5.8E-04 in
0.100 in	0.101 in	0.0010 in	± 0.001 in	±	5.8E-04 in
0.125 in	0.124 in	-0.0010 in	± 0.001 in	±	5.8E-04 in
0.200 in	0.199 in	-0.0010 in	± 0.001 in	±	5.9E-04 in
0.250 in	0.249 in	-0.0010 in	± 0.001 in	±	6.0E-04 in
0.300 in	0.299 in	-0.0010 in	± 0.001 in	±	6.1E-04 in
0.500 in	0.499 in	-0.0010 in	± 0.001 in	±	6.7E-04 in