

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	1 mm	1.04 mm	40 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	2 mm	1.53 mm	30 µm	± 30 µm	± 5.8E+00 µm
Profundidad / (Depth)	2 mm	2.03 mm	30 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	3 mm	2.97 mm	-30 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	5 mm	4.97 mm	-30 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	10 mm	9.97 mm	-30 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	15 mm	14.97 mm	-30 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	20 mm	19.97 mm	-30 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	25 mm	24.96 mm	-40 µm	± 30 µm	± 6.1E+00 µm