

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0625 in	0.062 in	-0.0005 in	± 0.002 in	±	5.8E-04 in
0.100 in	0.102 in	0.002 in	± 0.002 in	±	5.8E-04 in
0.125 in	0.123 in	-0.002 in	± 0.002 in	±	5.8E-04 in
0.200 in	0.202 in	0.002 in	± 0.002 in	±	5.9E-04 in
0.250 in	0.248 in	-0.002 in	± 0.002 in	±	6.0E-04 in
0.300 in	0.298 in	-0.002 in	± 0.002 in	±	6.1E-04 in
0.500 in	0.497 in	-0.003 in	± 0.002 in	±	6.7E-04 in
0.750 in	0.747 in	-0.003 in	± 0.002 in	±	7.8E-04 in
1.000 in	0.998 in	-0.002 in	± 0.002 in	±	9.1E-04 in