Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Exterior / (Exterior)	1.000 in	1.001 in	0.001 in	± 0.0019685 in	± 1.8E-03 in
Exterior / (Exterior)	4.000 in	4.001 in	0.001 in	$\pm 0.0027559 \text{ in}$	± 1.8E-03 in
Exterior / (Exterior)	8.000 in	8.001 in	0.001 in	$\pm 0.0031496 \text{ in}$	$\pm$ 1.8E-03 in
Exterior / (Exterior)	10.000 in	10.002 in	0.002 in	$\pm 0.0031496 \text{ in}$	$\pm$ 1.8E-03 in
Exterior / (Exterior)	12.000 in	12.002 in	0.002 in	$\pm 0.0035433$ in	$\pm$ 1.8E-03 in
Interior / (Interior)	4.000 in	4.001 in	0.001 in	$\pm 0.0027559 \text{ in}$	± 9.0E-04 in
Interior / (Interior)	12.000 in	12.002 in	0.002 in	$\pm 0.0035433$ in	± 9.0E-04 in
Profundidad / (Depth)	4.000 in	4.002 in	0.002 in	$\pm 0.0027559 \text{ in}$	± 1.1E-03 in
Profundidad / (Depth)	12.000 in	12.002 in	0.002 in	$\pm 0.0035433$ in	± 1.1E-03 in