Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Exterior / (Exterior)	1.2 in	1.2001 in	0.0001 in	± 0.0012 in	± 5.7E-02 in
Exterior / (Exterior)	2.4 in	2.4000 in	0.0000 in	± 0.0012 in	± 1.9E-04 in
Exterior / (Exterior)	3.6 in	3.6000 in	0.0000 in	± 0.0012 in	± 1.9E-04 in
Exterior / (Exterior)	4.8 in	4.7999 in	-0.0001 in	± 0.0012 in	± 1.9E-04 in
Exterior / (Exterior)	6 in	5.9999 in	-0.0001 in	± 0.0012 in	± 1.9E-04 in
Interior / (Interior)	0.6 in	0.5999 in	-0.0001 in	± 0.0012 in	\pm 2.3E-04 in
Interior / (Interior)	5.4 in	5.4002 in	0.0002 in	± 0.0012 in	\pm 2.4E-04 in
Profundidad / (Depth)	0.6 in	0.6001 in	0.0001 in	± 0.0012 in	± 1.7E-04 in
Profundidad / (Depth)	5.4 in	5.4001 in	0.0001 in	± 0.0012 in	± 1.7E-04 in