

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	1.20 in	1.2000 in	0 in	± 0.0012 in	± 2.9E-04 in
Profundidad / (Depth)	3.60 in	3.6000 in	0 in	± 0.0012 in	± 2.9E-04 in
Profundidad / (Depth)	4.5 in	4.5000 in	0 in	± 0.0012 in	± 2.9E-04 in
Profundidad / (Depth)	6 in	6.0000 in	0 in	± 0.0012 in	± 2.9E-04 in
Profundidad / (Depth)	7 in	7.2000 in	0 in	± 0.0012 in	± 2.9E-04 in
Profundidad / (Depth)	8 in	8.4000 in	0 in	± 0.0016 in	± 2.9E-04 in
Profundidad / (Depth)	10 in	9.6000 in	0 in	± 0.0016 in	± 2.9E-04 in
Profundidad / (Depth)	10 in	10.0000 in	0 in	± 0.0016 in	± 2.9E-04 in
Profundidad / (Depth)	12 in	12.0000 in	0 in	± 0.0016 in	± 2.9E-04 in