

Tipo de Medida (<i>Measurement Type</i>)	Patrón (<i>Standard</i>)	UBP (<i>UT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)
Profundidad / (Depth)	30 mm	30.00 mm	0 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	60.0 mm	60.00 mm	0 μm	± 30 μm	± 5.8E+00 μm
Profundidad / (Depth)	90 mm	90.00 mm	0 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	120 mm	120.00 mm	0 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	150 mm	150.00 mm	0 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	180 mm	180.00 mm	0 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	210 mm	210.00 mm	0 μm	± 40 μm	± 6.1E+00 μm
Profundidad / (Depth)	270 mm	270.00 mm	0 μm	± 40 μm	± 6.1E+00 μm
Profundidad / (Depth)	300 mm	300.00 mm	0 μm	± 40 μm	± 6.1E+00 μm