

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
23.2 μm	23.6 μm	0.4 μm	± 0.70 μm	±	5.8E-01 μm
51.1 μm	52.1 μm	1.0 μm	± 1.56 μm	±	5.8E-01 μm
125.1 μm	127 μm	1.9 μm	± 3.80 μm	±	5.8E-01 μm
478.0 μm	483 μm	5.0 μm	± 14.50 μm	±	5.8E-01 μm
963 μm	970 μm	7.0 μm	± 29.10 μm	±	6.0E-01 μm