

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.2 μm	23.0 μm	-0.2 μm	± 0.7 μm	±	5.8E-01 μm
51.1 μm	52.0 μm	0.9 μm	± 1.5 μm	±	5.8E-01 μm
125.1 μm	127 μm	1.9 μm	± 3.8 μm	±	5.8E-01 μm
478.0 μm	483 μm	5.0 μm	± 14.5 μm	±	5.8E-01 μm
963 μm	970 μm	7.0 μm	± 29.1 μm	±	6.0E-01 μm