

| Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |              |
|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|--------------|
| 49.98 mmHg                 | 50 mmHg            | 0.02 mmHg              | ± 3 mmHg                      | ±                                    | 6.6E-01 mmHg |
| 99.97 mmHg                 | 100 mmHg           | 0.03 mmHg              | ± 3 mmHg                      | ±                                    | 6.5E-01 mmHg |
| 149.98 mmHg                | 150 mmHg           | 0.02 mmHg              | ± 3 mmHg                      | ±                                    | 6.6E-01 mmHg |
| 199.98 mmHg                | 200 mmHg           | 0.02 mmHg              | ± 3 mmHg                      | ±                                    | 6.8E-01 mmHg |
| 249.97 mmHg                | 250 mmHg           | 0.03 mmHg              | ± 3 mmHg                      | ±                                    | 7.5E-01 mmHg |
| 299.97 mmHg                | 300 mmHg           | 0.03 mmHg              | ± 3 mmHg                      | ±                                    | 7.5E-01 mmHg |