

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)	
1.00 mm	1.001 mm	1 μm	± 8 μm	±	5.8E-01 μm
1.25 mm	1.251 mm	1 μm	± 8 μm	±	5.8E-01 μm
1.50 mm	1.501 mm	1 μm	± 8 μm	±	5.8E-01 μm
2.00 mm	2.001 mm	1 μm	± 8 μm	±	5.8E-01 μm
3.00 mm	3.001 mm	1 μm	± 8 μm	±	5.8E-01 μm
5.00 mm	5.001 mm	1 μm	± 8 μm	±	5.8E-01 μm
10.00 mm	10.001 mm	1 μm	± 8 μm	±	5.8E-01 μm
15.00 mm	15.001 mm	1 μm	± 8 μm	±	5.8E-01 μm
18.00 mm	18.001 mm	1 μm	± 8 μm	±	5.9E-01 μm
20.00 mm	20.002 mm	2 μm	± 8 μm	±	5.9E-01 μm
25.00 mm	25.002 mm	2 μm	± 8 μm	±	5.9E-01 μm