Patr	ón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumb	ore (Uncertainty)
0	100 in	0.100 in	0.000 in	± 0.001 in	±	7.1E-04 in
0	125 in	0.125 in	0.000 in	± 0.001 in	±	7.1E-04 in
0	200 in	0.200 in	0.000 in	± 0.001 in	±	7.2E-04 in
0	250 in	0.250 in	0.000 in	± 0.001 in	±	7.2E-04 in
0	300 in	0.300 in	0.000 in	± 0.001 in	±	7.3E-04 in
0	350 in	0.350 in	0.000 in	± 0.001 in	±	7.4E-04 in
0	400 in	0.400 in	0.000 in	$\pm 0.001 in$	±	7.6E-04 in