

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.000 mm	1.000 mm	0.000 µm	± 5 µm	±	5.8E-01 µm
1.250 mm	1.251 mm	1.000 µm	± 5 µm	±	5.8E-01 µm
1.500 mm	1.500 mm	0.000 µm	± 5 µm	±	5.8E-01 µm
2.000 mm	2.001 mm	1.000 µm	± 5 µm	±	5.8E-01 µm
3.000 mm	3.000 mm	0.000 µm	± 5 µm	±	5.8E-01 µm
5.000 mm	5.000 mm	0.000 µm	± 5 µm	±	5.8E-01 µm
6.000 mm	6.001 mm	1.000 µm	± 5 µm	±	5.8E-01 µm
8.000 mm	8.001 mm	1.000 µm	± 5 µm	±	5.8E-01 µm
10.000 mm	10.001 mm	1.000 µm	± 5 µm	±	5.8E-01 µm