

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.2 μm	23.3 μm	0.1 μm	± 2.5 μm	±	5.8E-02 μm
51.1 μm	51.3 μm	0.2 μm	± 2.5 μm	±	5.8E-02 μm
125.1 μm	126 μm	0.7 μm	± 3.8 μm	±	6.1E-02 μm
478.0 μm	479 μm	1.0 μm	± 14.3 μm	±	9.3E-02 μm
963 μm	966 μm	3.0 μm	± 28.6 μm	±	1.6E-01 μm