

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.3	0.29964	0.360	± 0.7	±	4.3E-01 mm						
0 a 0.6	0.59954	0.460	± 0.8	±	4.3E-01 mm						
0 a 0.9	0.89959	0.410	± 0.9	±	4.3E-01 mm						
0 a 1.2	1.19965	0.350	± 1	±	4.8E-01 mm						
0 a 1.5	1.49981	0.190	± 1.2	±	4.8E-01 mm						
0 a 1.8	1.79958	0.420	± 1.3	±	4.8E-01 mm						
0 a 2.1	2.09947	0.530	± 1.4	±	4.8E-01 mm						
0 a 2.4	2.39973	0.270	± 1.5	±	4.8E-01 mm						
0 a 2.7	2.69933	0.670	± 1.6	±	4.8E-01 mm						
0 a 3	2.99926	0.740	± 1.8	±	4.8E-01 mm						